

F1 F2



205596

Main Fuses: T-10A-H-250V
 Type: GXDP-300
 Ser. No: GO1601942
 Manufactured: October 2016
 Line voltage: 100 - 120V~ or 220 - 240V~, 50/60Hz
 Input power: stand by: 26W, maximum: 1100W
 Line impedance (max): 0.5 ohm
 Lifting motor duty cycle: 6.25%, 25s ON, 400s OFF
 Max line fusing:
 16A @ 100 - 120 VAC, 10 A @ 220 - 240 VAC

IP20
 IEC 60601-1
 CE 0537

This product complies with DHHS 21 CFR Chapter I Subchapter J at the date of manufacture.
 Rx only

ETL CLASSIFIED
 CONFORMS TO UL STD 60601-1.
 CERTIFIED TO CSA
 STD C22.2 NO 601.1.

Intertek
 3159129

Manufactured for Gendex Dental Systems
 Manufacturer: PaloDEX Group Oy
 Nahkelantie 160
 FI-04300 TUUSULA, Finland

201710-14

CLASS 1 LASERPRODUCT
 APPAREIL A RAYONNEMENT
 LASER DE KLASSE 1
 LASER KLASSE 1
 KLASS 1 LASERAPPARAT
 LJOKAN 1 LASERLAITE
 IEC 60825-1:2007





